

Arete

Sample: 01-16-2024-44371

Sample Received: 01/16/2024;

Report Created: 03/04/2024; Expires: 01/17/2025

Juicee J

Plant, Flower - Uncured



**19.486 %**

Total THC

**0.180 %**

Δ-9 THC

**22.296 %**

Total Cannabinoids

**<LOQ %**

Total CBD

## Cannabinoids

(Testing Method: HPLC, CON-P-3000)

Date Tested: 01/16/2024

Complete

Analyte	LOD	LOQ	Mass	Mass	
	%	%	%	mg/g	
Δ-8-Tetrahydrocannabinol (Δ-8 THC)	0.0476	0.0714	ND	ND	
Δ-9-Tetrahydrocannabinol (Δ-9 THC)	0.0476	0.0714	0.180	1.800	
Δ-9-Tetrahydrocannabinolic Acid (THCA-A)	0.0476	0.0714	22.013	220.133	
Δ-9-Tetrahydrocannabiphlorol (Δ-9-THCP)	0.0476	0.0714	ND	ND	
Δ-9-Tetrahydrocannabivarin (Δ-9-THCV)	0.0476	0.0714	ND	ND	
Δ-9-Tetrahydrocannabivarinic Acid (Δ-9-THCVA)	0.0238	0.0714	<LOQ	<LOQ	
R-Δ-10-Tetrahydrocannabinol (R-Δ-10-THC)	0.0476	0.0714	ND	ND	
S-Δ-10-Tetrahydrocannabinol (S-Δ-10-THC)	0.0476	0.0714	ND	ND	
9R-Hexahydrocannabinol (9R-HHC)	0.0476	0.0714	ND	ND	
9S-Hexahydrocannabinol (9S-HHC)	0.0476	0.0714	ND	ND	
Tetrahydrocannabinol Acetate (THCO)	0.0476	0.0714	ND	ND	
Cannabidivarin (CBDV)	0.0476	0.0714	ND	ND	
Cannabidivarinic Acid (CBDVA)	0.0476	0.0714	ND	ND	
Cannabidiol (CBD)	0.0476	0.0714	ND	ND	
Cannabidiolic Acid (CBDA)	0.0238	0.0714	<LOQ	<LOQ	
Cannabigerol (CBG)	0.0238	0.0714	<LOQ	<LOQ	
Cannabigerolic Acid (CBGA)	0.0476	0.0714	0.103	1.029	
Cannabinol (CBN)	0.0476	0.0714	ND	ND	
Cannabinolic Acid (CBNA)	0.0476	0.0714	ND	ND	
Cannabichromene (CBC)	0.0476	0.0714	ND	ND	
Cannabichromenic Acid (CBCA)	0.0476	0.0714	<LOQ	<LOQ	
<b>Total</b>			<b>22.296</b>	<b>222.962</b>	

Total THC = THCA \* 0.877 + Δ-9-THC; Total CBD = CBDA \* 0.877 + CBD; LOQ = Limit of Quantitation; ND = Not Detected.

Total THC Measurement of Uncertainty: ± 0.050%

Total CBD Measurement of Uncertainty: ± 2.000%

THCO potency analysis does not designate quantitative specificity of Δ-8-THCO and Δ-9-THCO isomers

Amended report issued to reflect change in sample identification.



New Bloom Labs  
6121 Heritage Park Drive, A500  
Chattanooga, TN 37416  
(844) 837-8223  
TN DEA#: RN0563975  
ANAB Testing Laboratory (AT-2868): ISO/IEC  
17025:2017

*Natalie Siracusa*  
Natalie Siracusa  
Laboratory Director

Powered by  
reLIMS  
info@relims.com